



**BID CLARIFICATION MATRIX NO.2**

**FOR**

**SUPPLY, DELIVERY, INSTALLATION, CALIBRATION AND MAINTENANCE OF RAPID TEST TOOL KITS TO SUPPORT THE VALUE CHAIN'S COMPLIANCE WITH QUALITY STANDARDS & SPS REQUIREMENTS**

**TENDER NUMBER: PRQ20241014**

**19<sup>TH</sup> MARCH 2025**

## SECTION 1: QUERIES AND RESPONSES

These clarifications are for those queries sought by bidders in line with the RFP of the tender document:

S/No.	Query	Response	Status (closed/ open/ pending)
1.	<p>Some product parameters are not clear about what is needed. They only provide a range or category name.</p> <p>Item 5: Target Pesticides: Organophosphorus, carbamate, pyrethroid, neonicotinoid, and other common pesticides</p>	<p>For rapid test kits, the specifications provided are the right ones</p>	<b>Closed</b>
2.	<p>Must specify a pesticide ? One test do not fit all Pesticides. If the customer does not know, we can provide the details.</p>	<p>We need a rapid test kits so the specifics are correct</p>	<b>Closed</b>
3.	<p>Item 8: Rapid test kits for Aflatoxin M1 , intended for the quantitative testing of aflatoxin in Milk and Milk products</p>	<p>The kit is for the rapid testing and it is a qualitative test</p>	<b>Closed</b>
4.	<p>Quantitative testing must use Reader to work together , you need to buy a machine) Reader / Machine / AccuScan Gold</p>	<p>This is not required since the analysis is qualitative and not quantitative</p>	<b>Closed</b>

**TRADEMARK AFRICA**

**18<sup>TH</sup> MARCH 2025**

